

Certificate of Analysis

No. CA 22130

Product: Eichrom's Anion Exchange Resin

Lot #: 060822A8M

Date of Manufacture: 8-JUN-2022


Eichrom's Anion Exchange resin Lot 060822A8M has been manufactured according to Eichrom standard operating procedures. This resin has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	0.03
Wet Volume Capacity	≥ 1.2 meq/mL	1.4

Jill Bryant

Quality System Coordinator



June 8, 2022