

# Certificate of Analysis

No. CA22154

Product: Eichrom's LN-2 Resin  
Lot #: L2S062922  
Date of Manufacture: 29-JUN-2022

Eichrom's LN-2 Resin Lot # L2S062922 has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White Beads	Pass
Lu Peak Maximum	11 - 15 BV	14.9 BV
Yb Peak Maximum	6.0 - 9.0 BV	8.9 BV
Seperation Factor	$\geq 1.6$	1.7

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

July 6, 2022