

# Certificate of Analysis

No. CA22157


Product: Eichrom's MnO<sub>2</sub> Resin  
Lot #: MNA061422  
Date of Manufacture: 14-JUN-2022

Eichrom's MnO<sub>2</sub> Resin Lot # MNA061422 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (mean, 4 reps)	> 95%	106.8 %
Ba-133 in Load solution (mean, 4 reps)	<1.0%	0.0 %

  
\_\_\_\_\_  
Sarah McAlister  
Quality Manager

  
\_\_\_\_\_  
July 13, 2022