

Certificate of Analysis

No. CA 22140


Product: Eichrom's Nickel Resin
Lot #: NIA052522
Date of Manufacture: 25-JUN-2022

Eichrom's Nickel Resin Lot # NIA052522 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	101.3 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %

Jill Bryant
Quality System Coordinator



June 15, 2022