

Certificate of Analysis
No. CA22161

Product: Eichrom's UTEVA® Resin UT15-C20-S-3PA Columns
Lot #: UT22D
Date of Manufacture: 11-JUL-2022

Eichrom's UTEVA® Resin UT15-C20-S-3PA columns Lot # UT22D has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, UTS041222, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	>100	165.7
% ²³⁰ Th stripped in 5M HCl/0.05M oxalic acid		
Least acceptable mean (4 rep)	90.0	99.1
Max. standard deviation	10.0	1.5
% ²³⁰ Th stripped in 1M HCl		
Maximum acceptable mean (5 rep)	0.5	0.1
Max. standard deviation	1.0	0.0
% ²³³ U stripped in 1M HCl		
Least acceptable mean (5 rep)	90.0	97.7
Max. standard deviation	10.0	1.6
% ²³³ U stripped in 5M HCl/0.05M oxalic acid		
Maximum acceptable mean (5 rep)	1.0	0.4
Max. standard deviation	1.0	0.2



Sarah McAlister
Quality Manager

15 Jul 22
July 15, 2022