

Certificate of Analysis

No. CA 22213

Product: Eichrom's Cation Exchange Resin Cartridges; 50Wx8, 100-200 mesh, H+ Form

Lot #: C8MR22E

Date of Manufacture: 13-SEP-2022

Eichrom's Cation Exchange resin cartridges, Lot # C8MR22E has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, 060121C8M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Extractables	< 1 mg(residue) / g(resin)	0.46 mg/g
pH	> 4.0	pass
Wet Volume Capacity	min. 1.7 meq/mL	2.0 meq/mL (Per vendor cert)

Jill Bryant
Quality System Coordinator

September 14, 2022