

Certificate of Analysis
No. CA22199

Product: Eichrom's DGA-Normal Resin
Lot #: DNS082522
Date of Manufacture: 25-AUG-2022

Eichrom's DGA-Normal Resin Lot # DNS082522 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution ratio, Eu in 0.1M nitric acid	≥900	1341
% ²⁴¹ Am stripped in 0.1M HCl		
Least acceptable mean (4 rep)	97.0	98.6
Max. standard deviation	2.5	0.3
% ²⁴¹ Am stripped in 0.1M HNO ₃		
Maximum acceptable mean (4 rep)	1.0	0.2
Max. standard deviation	0.5	0.3
% ²³³ U stripped in 0.1M HNO ₃		
Least acceptable mean (4 rep)	97.0	98.7
Max. standard deviation	2.5	0.1
% ²³³ U stripped in 0.1M HCl		
Maximum acceptable mean (4 rep)	1.0	0.8
Max. standard deviation	0.5	0.1

Jill Bryant
Quality System Coordinator

August 30, 2022