

Certificate of Analysis

No. CA 22185

Product: Eichrom's Nickel Resin Cartridges

Lot #: NIAR22C

Date of Manufacture: 15-AUG-2022

Eichrom's Nickel Resin cartridge Lot NIAR22C has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, # NIA052522, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	101.3 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %


Jill Bryant
Quality System Coordinator

August 16, 2022