

# Certificate of Analysis


No. CA22239

Product: Eichrom's TEVA® Discs  
Lot #: TEFD22A  
Date of Manufacture: 19-OCT-2022

Eichrom's TEVA® Discs, Lot # TEFD22A has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	47mm, round	Pass
Tc-99 yield in pH 2 (100-200 mL/min)	90 – 110 %	97.0 %

  
Jill Bryant  
Quality System Coordinator

October 19, 2022