

**Certificate of Analysis**  
**No. CA 22247**

Product: Eichrom's Cation Exchange Resin; 50Wx4, 100-200 mesh, Hydrogen Form Cartridges  
Lot: C4MR22A  
Date of Manufacture: 8-NOV-2022

Eichrom's Cation Exchange resin cartridges, Lot C4MR22A, has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, # 060718C4M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

**ANALYSIS RESULTS**

Test Value	Specification	Result
Extractables	< 1 mg(residue) / g(resin)	0.72 mg/g
pH	> 4.0	pass
Wet Volume Capacity	min. 1.1 meq/mL	2.1 meq/mL

  
Jill Bryant  
Quality System Coordinator

November 8, 2022