

Certificate of Analysis

No. CA 22291

Product: Eichrom's Anion Exchange Resin 2ml Cartridges

Lot #: A8MR22C

Date of Manufacture: 14-DEC-2022

Eichrom's Anion Exchange resin 2mL cartridge Lot A8MR22C has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, 070522A8M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	0.03
Wet Volume Capacity	≥ 1.2 meq/mL	1.4


Jill Bryant
Quality System Coordinator

December 19, 2022