



eichrom
TECHNOLOGIES

A GCI COMPANY

Certificate of Analysis

No. CA 22270

Product: Eichrom's Pb Resin
Lot #: PBA110922
Date of Manufacture: 09-NOV-2022

Eichrom's Pb Resin Lot # PBA110922, has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% Lead in Load + Rinse	< 0.1 %	0.0 %
% Lead in Strip	> 99 %	100.0 %
Gravimetric Yield	80-120 %	91.3 %
Chemical Yield	80-120 %	100.6 %

Jill Bryant
Quality System Coordinator

November 30, 2022

