

Certificate of Analysis

No. CA 22295

Product: Eichrom's Sr Resin 2mL Cartridges
 Lot #: SRSR22I
 Date of Manufacture: 20-DEC-2022

Eichrom's Sr Resin cartridge Lot SRSR22I has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, SRS120722, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	≥ 100	110.2
% Sr stripped in 0.05M HNO ₃		
Least acceptable mean (5 rep)	85.0	97.5
Max. standard deviation	2.5	1.3
%Ca stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.15	0.00
Max. standard deviation	0.05	0.01
%Ba stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.75	0.16
Max. standard deviation	0.10	0.00
%Y stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.15	0.00
Max. standard deviation	0.05	0.00
%Y in 8M HNO ₃ load and rinse		
Least acceptable mean (5 rep)	99.0	100.0
Max. standard deviation	2.5	0.0


 Jill Bryant
 Quality System Coordinator

December 20, 2022