

# Certificate of Analysis


No. CA 23001

Product: Eichrom's Anion Exchange Resin  
Lot #: 010323A4M  
Date of Manufacture: 3-JAN-2023

Eichrom's Anion Exchange resin, Lot # 010323A4M has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	0.06
Wet Volume Capacity	≥ 1.0 meq/mL	1.2

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

January 6, 2023