

Certificate of Analysis


No. CA 23054

Product: Eichrom's 25mm Resolve™ PE Filters in Disposable Funnels
Lot #: RFDF032223F
Date of Manufacture: 22-MAR-2023

Eichrom's 25mm Resolve™ PE Filters in Disposable Funnels Lot RFDF032223F has been manufactured according to Eichrom standard operating procedures. The PE filter membrane used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	42.2
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	38.9
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0


Jill Bryant
Quality System Coordinator

April 4, 2023