

Certificate of Analysis

No. CA 23111

Product: Eichrom's Anion Exchange Resin
Lot #: 060123A4M
Date of Manufacture: 1-JUN-2023

Eichrom's Anion Exchange resin, Lot # 060123A4M has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	0.0
Wet Volume Capacity	≥ 1.0 meq/mL	1.2


Jill Bryant
Quality System Coordinator

June 2, 2023