

Certificate of Analysis

No. CA 23119

Product: Eichrom's Actinide Resin Cartridges

Lot #: ACSR23C

Date of Manufacture: 2-JUN-2023

Eichrom's Actinide Resin cartridge Lot ACSR23C has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, ACS040423, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.5 M HCl acid	≥ 300	1000


Jill Bryant
Quality System Coordinator

June 6, 2023