

Certificate of Analysis

No. CA23073

Product: Eichrom's MnO₂ Resin Cartridges
Lot #: MNAR23B
Date of Manufacture: 24-APR-2023

Eichrom's MnO₂ Resin cartridge Lot MNAR23B has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, MNA061422, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (mean, 4 reps)	> 95%	106.8 %
Ba-133 in Load solution (mean, 4 reps)	<1.0%	0.0 %

Jill Bryant
Quality System Coordinator



April 24, 2023