

# Certificate of Analysis


No. CA 23126


Product: Eichrom's Nickel Resin Columns  
Lot #: NI23A  
Date of Manufacture: 19-JUN-2023

Eichrom's Nickel Resin column Lot NI23A has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, NIA110722, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	101.0 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %

  
\_\_\_\_\_  
Sarah McAlister  
Quality Manager

  
\_\_\_\_\_  
June 20, 2023