

Certificate of Analysis


No. CA 23161

Product: Eichrom's Pb Resin QML Cartridges
Lot #: PBFR23A
Date of Manufacture: 27-JUL-2023

Eichrom's Pb Resin QML Cartridges Lot # PBFR23A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, PBF092822, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% Lead in Load + Rinse	< 0.1 %	0.0 %
% Lead in Strip	> 99 %	100.0 %
Gravimetric Yield	80-120 %	84.7 %
Chemical Yield	80-120 %	103.7 %


Jill Bryant
Quality System Coordinator

July 27, 2023