

Certificate of Analysis

No. CA 23076

Product: Eichrom's L grade Pre-filter Resin, > 150µm Cartridges

Lot #: PFLR23A


Date of Manufacture: 26-APR-2023

Eichrom's L grade Pre-filter Resin cartridge Lot PFLR23A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, PFL022321, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Flow Rate	> 0.80 mL/min	1.08

Jill Bryant
Quality System Coordinator



April 27, 2023