

Certificate of Analysis

No. CA 23135

Product: Eichrom's 25mm Resolve™ PE Filters
Lot #: RF070123
Date of Manufacture: 1-JUL-2023

Eichrom's 25mm Resolve™ Filters Lot # RF070123 has been manufactured according to Eichrom standard operating procedures. The PE filter membrane used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	42.2
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	38.9
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0


Jill Bryant
Quality System Coordinator

July 5, 2023