

Certificate of Analysis

No. CA 23141

Product: Eichrom's Sr Resin Cartridges
Lot #: SRSR23E
Date of Manufacture: 11-JUL-2023

Eichrom's Sr Resin cartridge Lot SRSR23E has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, SRS052323, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	≥ 100	103.8
% Sr stripped in 0.05M HNO ₃		
Least acceptable mean (5 rep)	85.0	95.5
Max. standard deviation	2.5	1.1
%Ca stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.15	0.03
Max. standard deviation	0.05	0.05
%Ba stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.75	0.03
Max. standard deviation	0.10	0.07
%Y stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.15	0.00
Max. standard deviation	0.05	0.00
%Y in 8M HNO ₃ load and rinse		
Least acceptable mean (5 rep)	99.0	100.0
Max. standard deviation	2.5	0.0


Jill Bryant
Quality System Coordinator

July 12, 2023