

Certificate of Analysis

No. CA 23169

Product: Eichrom's Anion Exchange Resin
 Lot #: 080123A8F
 Date of Manufacture: 1-AUG-2023

Eichrom's Anion Exchange resin Lot 080123A8F has been manufactured according to Eichrom standard operating procedures. This resin has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to Light Yellow	Pass
Extractables	< 1 mg (Residue) / g (Resin)	0.03
Wet Volume Capacity	≥ 1.2 meq/mL	1.5



Sarah McAlister
Quality Manager

August 2, 2023