

Certificate of Analysis

No. CA 23231

Product: Eichrom's Cation Exchange Resin; 50Wx8, 100-200 mesh, H+ Form
 Lot #: 092723C8M
 Date of Manufacture: 27-SEP-2023

Eichrom's Cation Exchange resin, Lot # 092723C8M has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Extractables	< 1 mg(residue) / g(resin)	0.12 mg/g
pH	> 4.0	pass
Wet Volume Capacity	min. 1.7 meq/mL	2.0 meq/mL (Per vendor cert)


 Jill Bryant
 Quality System Coordinator

September 28, 2023