

Certificate of Analysis

No. CA 23298

Product: Eichrom's Anion Exchange Resin

Lot #: 121023PA8M

Date of Manufacture: 10-DEC-2023

Eichrom's Anion Exchange resin Lot 121023PA8M has been manufactured according to Eichrom standard operating procedures. This resin has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	0.1
Wet Volume Capacity	≥ 1.2 meq/mL	1.6



Jill Bryant
Quality System Coordinator

December 14, 2023