

# Certificate of Analysis


No. CA 23285

Product: Eichrom's Anion Exchange Resin Cartridges  
Lot #: A4MR23A  
Date of Manufacture: 20-NOV-2023

Eichrom's Anion Exchange resin cartridge Lot A4MR23A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, # 091721A4M, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White to light Yellow	Pass
Extractables	< 1 mg (residue) / g (resin)	n.d
Wet Volume Capacity	≥ 1.0 meq/mL	1.2

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

November 20, 2023