

Certificate of Analysis

No. CA 23187

Product: Eichrom's Cation Exchange Resin; 50Wx8, 200-400 mesh, H⁺ Form Cartridges
Lot #: C8FR23E
Date of Manufacture: 11-AUG-2023

Eichrom's Cation Exchange resin cartridge Lot C8FR23E has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, 062723C8F, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

Test Value	Specification	Result
Extractables	< 1 mg(residue) / g(resin)	0.3 mg/g
pH	> 4.0	pass
Wet Volume Capacity	min. 1.7 meq/mL	2.0 meq/mL


Jill Bryant
Quality System Coordinator

August 15, 2023