

Certificate of Analysis

No. CA23265

Product: Eichrom's LN-2 Resin Cartridges
Lot #: L2SR23B
Date of Manufacture: 25-OCT-2023

Eichrom's LN-2 Resin cartridge Lot L2SR23B has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot L2S092523, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White Beads	Pass
Lu Peak Maximum	11 - 15 BV	13.96 BV
Yb Peak Maximum	6.0 - 9.0 BV	7.96 BV
Seperation Factor	≥ 1.6	1.82


Jill Bryant
Quality System Coordinator

October 26, 2023