

# Certificate of Analysis

No. CA23238

Product: Eichrom's LN Resin Columns  
Lot #: LN23B  
Date of Manufacture: 4-OCT-2023

Eichrom's LN Resin column Lot LN23B has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, LNA082423, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance Distribution Ratio, 0.1 N nitric acid	White beads >700	Pass 2621

  
\_\_\_\_\_  
Jill Bryant  
Quality System Coordinator

October 4, 2023