

# Certificate of Analysis

No. CA23175

Product: Eichrom's LN Resin Cartridges  
Lot #: LNSR23B  
Date of Manufacture: 3-AUG-2023

Eichrom's LN Resin cartridge Lot LNSR23B has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, LNS071723, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 0.1 N nitric acid	>700	2612



---

Sarah McAlister  
Quality Manager

August 4, 2023