



A GCI COMPANY

Certificate of Analysis

No. CA 24008

Product: Eichrom's MnO₂ Resin
Lot #: MNA011524
Date of Manufacture: 15-JAN-2024

Eichrom's MnO₂ Resin Lot # MNA011524 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Ba-133 recovery (mean, 4 reps)	> 95%	107.7 %
Ba-133 in Load solution (mean, 4 reps)	<1.0%	0.0 %

Jill Bryant
Quality System Coordinator

January 18, 2024