

Certificate of Analysis

No. CA 23203

Product: Eichrom's Nickel Resin Columns
Lot #: NI23B
Date of Manufacture: 29-AUG-2023

Eichrom's Nickel Resin column Lot NI23B has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, NIA110722, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	101.0 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %



Jill Bryant
Quality System Coordinator

August 29, 2023