



A GCI COMPANY

Certificate of Analysis

No. CA 23202

Product: Eichrom's Nickel Resin 5mL Columns
Lot #: NI23C
Date of Manufacture: 29-AUG-2023

Eichrom's Nickel Resin 5mL Columns Lot NI23C has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, NIA050223, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	101.8 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %

Jill Bryant
Quality System Coordinator

August 29, 2023

