

# Certificate of Analysis

No. CA 23282


Product: Eichrom's Nickel Resin Cartridges  
Lot #: NIAR23C  
Date of Manufacture: 20-NOV-2023

Eichrom's Nickel Resin cartridge Lot NIAR23C has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, NIA102723, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	103 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %

Jill Bryant  
Quality Analyst



November 20, 2023