



Certificate of Analysis

No. CA 23301

Product: Eichrom's Nickel Resin Cartridges
Lot #: NIAR23D
Date of Manufacture: 14-DEC-2023

Eichrom's Nickel Resin cartridge Lot NIAR23D has been manufactured according to Eichrom's standard operating procedures. The resin used to manufacture this lot, NIA102723, has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Ave. % Ni Recovery	80-105%	103 %
Ave. % Ni in Load/Rinse	<1%	0.2 %
Ave. % Ni in Strip	>99%	99.8 %

Jill Bryant
Quality System Coordinator

December 18, 2023