

Certificate of Analysis


No. CA 23189

Product: Eichrom's Pb Resin
Lot #: PBA080223
Date of Manufacture: 2-AUG-2023

Eichrom's Pb Resin Lot # PBA080223, has been manufactured according to Eichrom standard operating procedures. This lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
% Lead in Load + Rinse	< 0.1 %	0.0 %
% Lead in Strip	> 99 %	100.0 %
Gravimetric Yield	80-120 %	82.1 %
Chemical Yield	80-120 %	103.2 %



Jose Munoz
Quality Analyst

August 17, 2023