

Certificate of Analysis

No. CA 23262

Product: Eichrom's L grade Pre-filter Resin, > 150µm Cartridges
Lot #: PFLR23D
Date of Manufacture: 23-OCT-2023

Eichrom's L grade Pre-filter Resin cartridge Lot PFLR23D has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, PFL022321, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Flow Rate	> 0.80 mL/min	1.08



Jill Bryant
Quality System Coordinator

October 24, 2023