

# Certificate of Analysis

No. CA 23228

Product: Eichrom's Pre-filter Resin, 50-100 µm, HML Cartridges  
Lot #: PFSR23C  
Date of Manufacture: 18-SEP-2023

Eichrom's S grade Pre-filter Resin cartridge Lot PFSR23C has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, YY00M37012, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Flow Rate	0.25-0.40 mL/min	0.30

  
Jose Munoz  
Quality Analyst

September 27, 2023