



Certificate of Analysis
No. CA 23280

Product: Eichrom's RE Resin Cartridges
Lot #: RESR23C
Date of Manufacture: 6-NOV-2023

Eichrom's RE Resin cartridge Lot RESR23C has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, # RES051222, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, 3.1 N nitric acid	> 125	150.0

Jose Munoz
Quality Analyst

November 8, 2023