

Certificate of Analysis

No. CA 24007

Product: Eichrom's 25mm Resolve™ PE Filters
 Lot #: RF011024
 Date of Manufacture: 10-JAN-2024

Eichrom's 25mm Resolve™ Filters Lot # RF011024 has been manufactured according to Eichrom standard operating procedures. The PE filter membrane used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	43.7
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	41.1
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0


 Jill Bryant
 Quality System Coordinator

January 18, 2024