

# Certificate of Analysis

No. CA 23173

Product: Eichrom's 25mm Resolve™ PE Filters  
Lot #: RF072423  
Date of Manufacture: 24-JUL-2023

Eichrom's 25mm Resolve™ Filters Lot # RF072423 has been manufactured according to Eichrom standard operating procedures. The PE filter membrane used to manufacture this lot has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets or exceeds all Eichrom quality specifications.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Am-241 FWHM avg keV (8 replicates)	<50.0	42.6
Am-241 FWHM values above 60 keV	0	0
Am-241 FWHM values between 50-60 keV	2 or fewer	0
Am-243 FWHM avg keV (8 replicates)	<50.0	39.6
Am-243 FWHM values above 60 keV	0	0
Am-243 FWHM values between 50-60 keV	2 or fewer	0



Sarah McAlister  
Quality Manager

August 2, 2023