



**eichrom**  
TECHNOLOGIES

A GCI COMPANY

# Certificate of Analysis

No. CA 24042

Product: SCX Cartridges, QML  
Part No: SCXQML-R100  
Lot No: SCXR24B  
Manufactured Date: 22-FEB-2024

Eichrom's SCX Cartridge Lot # SCXR24B has been manufactured according to Eichrom's standard operating procedures. The silica used to manufacture this lot, SCX020824, has been tested in accordance with Eichrom's quality assurance procedures and purity testing. This lot of product meets all Eichrom quality specifications.

No products of animal or human origin are used in the manufacture of this product.

## ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Fe Content (µg per cartridge)	<0.6	0.2
Recovery of Ga (%)	>98.0	99.7

Jill Bryant  
Quality System Coordinator

February 28, 2024