

Certificate of Analysis

No. CA 23252

Product: Eichrom's Sr Resin
Lot #: SRS101723
Date of Manufacture: 17-OCT-2023

Eichrom's Sr Resin Lot # SRS101723 has been manufactured according to Eichrom's standard operating procedures. This lot has been tested in accordance with Eichrom's quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance	White beads	Pass
Distribution Ratio, Sr in 3.1 N nitric acid	≥ 100	110
% Sr stripped in 0.05M HNO ₃		
Least acceptable mean (5 rep)	85.0	94.9
Max. standard deviation	2.5	2.0
%Ca stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.15	0.01
Max. standard deviation	0.05	0.02
%Ba stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.75	0.06
Max. standard deviation	0.10	0.08
%Y stripped in 0.05M HNO ₃		
Maximum acceptable mean (5 rep)	0.15	0.00
Max. standard deviation	0.05	0.00
%Y in 8M HNO ₃ load and rinse		
Least acceptable mean (5 rep)	99.0	100.0
Max. standard deviation	2.5	0.0

Jill Bryant
Quality System Coordinator

October 19, 2023