

Certificate of Analysis

No. CA 24124

Product: Eichrom's LN Resin Cartridges
Lot #: LNSR24B
Date of Manufacture: 29-MAY-2024

Eichrom's LN Resin cartridge Lot LNSR24A has been manufactured according to Eichrom standard operating procedures. The resin used to manufacture this lot, LNS032824, has been tested in accordance with Eichrom quality assurance policy and procedures. This lot of product meets all Eichrom quality specifications.

ANALYSIS RESULTS

<u>Test Value</u>	<u>Specification</u>	<u>Result</u>
Appearance Distribution Ratio, 0.1 N nitric acid	White beads >700	Pass 2676



Jose Munoz
Quality Analyst

May 30, 2024